

Fig. 1a

Basic LS-APGD Source Operation




Fig. 1b

Proposed Implementation of LS-APGD with Microfluidic Devices




Fig. 1c




Fig. 2a






Fig. 2b








Fig. 4a




Fig. 5





Fig. 6

Fig. 7







Fig. 9



Analytical response functions and limits of detection for the LS-APGD device. Solution flow rate = 1 mL/min, electrolyte pH = 1, inter-electrode gap = 1 mm, injection volume = 5 μ L.

| Element | wavelength (nm) | peak height eqn. R^2 | peak area eqn. R^2 | LOD ppm (ng) |
|---------|--------------------|---------------------------|---------------------------|-----------------|
| Na | 589.0 | $y=0.421x + 42.8$ 0.9859 | $y=15.81x + 978.6$ 0.9784 | 12 (60) |
| Fe | 248.3 | $y=1.06x - 102.1$ 0.9365 | $y=45.80x - 6649$ 0.909 | 12 (60) |
| Pb | 405.8 | $y=1.18x - 10.45$ 0.977 | $y=16.16x - 419.7$ 0.9298 | 14 (70) |

FIG. 10




Fig. 11




Fig. 12





FIG. 13



Selenoamino Acid Separation




Fig. 15